

Tinytag Ultra 2 Temperature Logger with Integral Sensor and for Thermistor Probe (-40 to +85 °C/-40 to +125 °C)

TGU-4510

Issue 11

9th August 2019
E&OE

Tinytag Ultra 2 data loggers are ideally suited to monitor interior applications where there is little or no moisture.

Tinytag Ultra 2 data loggers have a high reading accuracy and resolution, large memories, a fast offload speed and a low battery monitor.


The TGU-4510 has an integral sensor and a connection for an external probe (not included) that makes the unit ideal for comparing ambient and process conditions.

Popular Applications

- Office and housing monitoring
- Pharmaceutical manufacture
- Dry food storage
- Museum display and repository
- Incubators

Features

- Dual channel temperature recorder
- 32,000 reading capacity
- High accuracy
- High reading resolution
- Fast data offload
- Splash-proof case
- Low battery monitor
- User-replaceable battery



For pricing or any further information, please contact Omni Instruments Ltd.



Features


Total Reading Capacity	32,000 readings
Memory type	Non Volatile
Trigger Start	Magnetic Switch (from SN 602211)
Delayed Start	Relative / Absolute (up to 45 days)
Stop Options	When full After n Readings Never (overwrite oldest data)
Reading Types	Actual, Min, Max
Logging Interval	1 sec to 10 days
Offload	While stopped or when logging in minutes mode
Alarms	2 fully programmable; latching

Reading Specification

Internal Temperature

Reading Range	-40 °C to +85 °C (-40 °F to +185 °F)
Sensor Type	10K NTC Thermistor (Internally mounted)
Response Time	20 mins to 90% FSD in moving air
Reading Resolution	0.01 °C or better

Accuracy




External Temperature

Reading Range	-40 °C to +125 °C (-40 °F to +257 °F)
Sensor Type	10K NTC Thermistor (External probe)
Reading Resolution	0.02 °C or better
Temperature Stability	±0.01 °C/°C change from 25 °C

Accuracy

The graph below is the accuracy of the TGU-4510 when used



Physical Specification

IP Rating	IP54 splash proof (see notes)
Operational Range*	-40 °C to +85 °C (-40 °F to +185 °F)
Case Dimensions	
Height	72mm / 2.83"
Width	60mm / 2.36"
Depth	33mm / 1.30"
Weight	55g / 1.94oz

*The Operational Range indicates the physical limits to which the unit can be exposed, not the reading range over which it will record.

Notes

The battery fitted in this product is a single cell containing less than 1g of lithium and meets the requirements of the UN Manual of Tests and Criteria, Part III, Subsection 38.3.

Recommended Battery Types	SAFT LS14250, Tekcell SBAA02P or Eve ER14250
----------------------------------	--

The logger will operate with other 1/2AA 3.6V Lithium batteries but performance cannot be guaranteed.

Replacement Interval	Annually
-----------------------------	----------

Before replacing the battery the data logger must be stopped.

When replacing the battery, wait at least one minute after removing the old battery before fitting the new one.

Data stored on the logger will be retained after a battery is replaced.

If used at low temperatures the data logger should be allowed to warm to room temperature before it is opened to avoid condensation forming inside the unit.

The IP54 rating is valid only when the unit's connector cap and probe are fitted and the unit is orientated with its hanging tab uppermost.

Trigger Start

The trigger start option allows a unit to be set up as required and then started at a later time with a magnet. The position of the unit's trigger start switch is indicated by the ••• label on the back of the logger. When the "Wait until trigger event" option is selected in the Tinytag Explorer software, the green LED on the unit will flash once every eight seconds, indicating that the unit is waiting to log. When a magnet passed over the label, the green LED will light briefly to indicate that the unit has been activated. Once activated, the green LED will flash every four seconds to indicate that the logger is recording.

Calibration

This unit is configured to meet Gemini's quoted specification during its manufacture.

We recommend that the calibration of this unit should be checked annually against a calibrated reference meter.

A certificate of calibration, traceable to a national standard, can be supplied for an additional charge either at the point of purchase, or if the unit is returned for a service calibration.

Approvals

Gemini Data Loggers (UK) Ltd. operates a Business Management System which conforms to ISO 9001 and ISO 14001.



For pricing or any further information, please contact Omni Instruments Ltd.



Required and Related Products

To use this data logger you will also require a probe:

PB-5001-1M5/3M/5M: Standard Thermistor Probe or
PB-5002-1M5/3M/5M: Fast Response Thermistor Probe or
PB-5003-1M5/3M/5M: Surface Thermistor Probe

(Other types of probes are available, please see
www.tinytag.info for further details).

The following piece of software:

SWCD-0040: Tinytag Explorer software

and a

CAB-0007-USB: Tinytag Ultra/Plus/View USB Download Cable

The SWCD-0040 software and CAB-0007-USB cable can be
ordered together in a pack using the part number SWPK-7-USB.

Further Related Products

SER-9500: Tinytag Data Logger Service Kit

ACS-6000: Trigger Start Magnet

data sheet

For pricing or any further information, please contact Omni Instruments Ltd.