



Tinytag View 2 Temperature/Relative Humidity Logger (-25 to +50 ℃/0 to 100% RH) Grey Case

TV-4501

Issue 8 9th August 2019 E&OE Tinytag View 2s are a range of data loggers with displays. Featuring high reading accuracy and resolution, large memories, a fast offload speed and a low battery monitor, these units are housed in splash-proof (IP65 rated) cases.

The TV-4501 is a self contained temperature and relative humidity data logger. This unit features a coated RH sensor that has good resistance to moisture and condensation, ensuring measurement reliability in applications where a visual display of temperature and humidity is required in addition to data logging.

This unit is an unobtrusive grey colour for use in museums and art galleries.

Popular Applications

- Museums and art galleries
- · Document and archive monitoring
- · Environmental monitoring



Features

- Temperature and relative humidity recorder
- · LCD display of current readings
- 30,000 reading capacity
- High accuracy
- · High reading resolution
- Fast data offload
- Splash-proof case
- Low battery monitor
- User-replaceable battery













For pricing or any further information, please contact Omni Instruments Ltd.



Contact Details: Tel: +44 1382 443000 Email: info@omni.uk.com Mailing Address: Unit 1, 14 Nobel Road, Wester Gourdie Industrial Estate, Dundee, DD2 4UH.

Website: www.omniinstruments.co.uk





Tinytag View 2 Temperature/Relative Humidity Logger (-25 to +50 °C/0 to 100% RH) Grey Case

Issue 8: 9th August 2019 (E&OE)



Features

Total Reading Capacity 30,000 readings Memory type Non Volatile Display 4 digits + indicators **Display Modes** ℃ or ℉ / %RH **Display Refresh Rate** Every 2 seconds

> (alternating temperature/humidity) Magnetic Switch

Trigger Start **Delayed Start** Relative / Absolute (up to 45 days) **Stop Options** When full

After n Readings Never (overwrite oldest data)

Reading Types Actual, Min, Max Logging Interval Offload

1 sec to 10 days While stopped or when logging in minutes

mode


2 fully programmable; latchable Alarms

Reading Specification

Temperature

Reading Range -25 °C to +50 °C (-13 °F to 122 °F) Sensor Type 10K NTC Thermistor (Integral) 10 mins to 90% FSD in moving air Response Time Logger Resolution 0.02℃ or better **Display Resolution** 0.1 °C or 0.1 °F

Logger Accuracy



Relative Humidity


Reading Range 0 to 100% RH Sensor Type Capacitive (Integral) ±3.0% RH at 25 °C / 77 °F Accuracy Reading Resolution Better than 0.3% RH 40 seconds to 90% FSD (current Response Time data loggers, from SN 612487)

Display Resolution 0.1% RH

RH Sensor Working Range

The working range for the RH sensor is shown in terms of relative humidity / temperature limits.

Although the sensor will not fail beyond these limits, the accuracy will deteriorate.



Physical Specification

IP65 splash proof (see notes) Operational Range* -25 °C to +70 °C

Case Dimensions

Diameter 60mm / 2.36" Length 90mm / 3.54" Width 65mm / 2.56" Depth 35mm / 1.38 Weight 85q / 3oz

*The Operational Range indicates the physical limits to which the unit can be exposed, not the reading range over which it will record.

Notes

The battery fitted in this product is a single cell containing less that 1g of lithium and meets the requirements of the UN Manual of Tests and Criteria, Part III, Subsection 38.3.

Recommended Battery Types SAFT LS14250,

Tekcell SBAA02P or Eve ER14250

The logger will operate with other ½AA 3.6V Lithium batteries but performance cannot be guaranteed.

Replacement Interval Annually

Before replacing the battery the data logger must be stopped.

After removing an old battery from a logger, wait five minutes before inserting the new one

Data stored on the logger will be retained after a battery is

The clarity of the display may change at extremes of temperature.

If used at low temperatures the data logger should be allowed to warm to room temperature before it is opened to avoid condensation forming inside the unit.

The IP65 rating is valid only when the unit's connector cap is

The coated sensor used on this unit (current product, SN 612487 and above) provides good protection from moisture and condensation, but in some cases - where the sensor becomes saturated - readings may become unpredictable. Once the sensor has dried out, and provided no residue is left behind, the unit should return to normal reading within 30 minutes

Any dust, salts or residue that is allowed to build up on the RH sensor will affect the unit's reading accuracy.

The sensor may be cleaned with de-ionised water but not with pure isopropanol or abrasive detergents, as these may damage the coating on the sensor and effect its accuracy.

The RH sensor will resist small amounts of the following chemicals: formaldehyde, ammonia, carbon monoxide, sulphur dioxide, ethylene oxide, hydrogen chloride, hydrogen fluoride, hydrogen peroxide, nitrogen dioxide, methyl chloride, chlorine, freon, methanol, ethanol, isopropanol and ozone. It also offers resistance to ultraviolet rays.

For pricing or any further information, please contact Omni Instruments Ltd.



Contact Details:

Tel: +44 1382 443000 Email: info@omni.uk.com Mailing Address: Unit 1, 14 Nobel Road, Wester Gourdie Industrial Estate, Dundee, DD2 4UH.

Website: www.omniinstruments.co.uk





Tinytag View 2 Temperature/Relative Humidity Logger (-25 to +50 ℃/0 to 100% RH) Grey Case

TV-4501

Issue 8: 9th August 2019 (E&OE)



Calibration

This unit is configured to meet Gemini's quoted accuracy specification during its manufacture.

We recommend that the calibration of this unit should be checked every six months against a calibrated reference meter.

A certificate of calibration, traceable to a national standard, can be supplied for an additional charge either at the point of purchase, or if the unit is returned for a service calibration.

Approvals

Gemini Data Loggers (UK) Ltd. operates a Business Management System which conforms to ISO 9001 and ISO 14001.



Required and Related Products

To use this data logger you will require the following software:

SWCD-0040: Tinytag Explorer software

and a

CAB-0007-USB: Tinytag Ultra/Plus/View USB Download Cable

The SWCD-0040 software and CAB-0007-USB cable can be ordered together in a pack using the part number SWPK-7-USB.

Further Related Products

SER-9500: Tinytag Data Logger Service Kit ACS-6000: Trigger Start Magnet

For pricing or any further information, please contact Omni Instruments Ltd.



Contact Details: Tel: +44 1382 443000 Email: info@omni.uk.com

tails: Mailing Address: Unit 1, 14 Nobel Road, 2 443000 Wester Gourdie Industrial Estate, Dundee, DD2 4UH.

Website: www.omniinstruments.co.uk